

PTCalibrationHistory

Maint Record #	Sensor ID	Slope	Calibration Offset	Std Error	Calibration Date
400295	4583	0.0057	-0.0536	0.0194	12/14/2010 1:38:29 PM
205192	1484	0.0113	0.0447	0.0108	9/8/2010 1:44:36 PM
400282	4403	0.0109	-0.063	0.0143	9/2/2010 1:09:22 PM
400280	4583	0.0058	-0.1231	0.0342	9/2/2010 11:52:35 AM
400277	803	0.01	0.0162	0.005	9/1/2010 1:40:20 PM
205150	713	0.0115	-0.0322	0.0075	8/26/2010 10:14:02 AM
205135	953	0.0109	0.1811	0.0774	8/18/2010 2:26:44 PM
205134	403	0.0114	0.0357	0.0094	8/18/2010 12:23:56 PM
400262	803	0.0102	-0.412	0.116	8/10/2010 3:44:14 PM
205074	4583	0.0059	0.0083	0.0025	7/30/2010 11:24:16 AM
205069	4383	0.0104	-0.0607	0.0164	7/29/2010 12:40:28 PM
205035	1484	0.0113	0.5111	0.123	7/21/2010 9:48:43 AM
400225	1003	0.0112	0.0122	0.006	7/19/2010 1:00:29 PM
205019	813	0.0113	0.0027	0.0014	7/14/2010 1:04:57 PM
205018	1803	0.0116	0.0377	0.009	7/14/2010 11:37:55 AM
205015	863	0.0115	-0.1157	0.0285	7/8/2010 2:11:27 PM
205006	323	0.0113	0.0259	0.0069	7/1/2010 12:43:13 PM
204984	4383	0.0101	-0.2352	0.0677	6/22/2010 10:49:30 AM
204915	323	0.0112	-0.2097	0.0736	5/4/2010 1:35:05 PM
204913	1723	0.0114	-0.0052	0.0051	5/4/2010 10:56:01 AM
400217	4583	0.0058	-0.0913	0.0134	4/15/2010 12:05:01 PM
204906	1303	0.0089	0.0924	0.0154	4/12/2010 1:33:40 PM
204904	1503	0.012	0.0731	0.017	4/12/2010 11:47:29 AM
204901	413	0.0114	0.3607	0.0861	4/9/2010 11:26:31 AM
204899	953	0.0107	0.0502	0.0165	4/8/2010 12:14:15 PM
400201	943	0.0119	0.1024	0.026	3/5/2010 12:28:38 PM
400198	953	0.0107	-0.002	0.0031	3/4/2010 2:47:40 PM
400190	813	0.0111	0.2946	0.0798	3/3/2010 2:09:41 PM
204863	1303	0.0114	-0.0241	0.0424	3/2/2010 12:51:16 PM
400181	433	0.0117	0.0498	0.0334	2/26/2010 2:42:37 PM
204858	123	0.0111	0.0098	0.0039	2/26/2010 1:49:23 PM
400179	403	0.0103	0.0061	0.0025	2/26/2010 11:36:44 AM
400175	1503	0.0111	-0.0354	0.0085	2/25/2010 12:53:25 PM
400172	1613	0.0113	-0.5477	0.1842	2/18/2010 1:08:29 PM